

# Technical Data Sheet

Doc No – Ab/TDS/CT-013

Rev No : 06 Date 02/2022-23

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**Rack for Micro Centrifuge Tube, PC**

**Valid for Catalog No. P20201, P20202, P20203, P20204**

### **1. Description/Specification**

#### **1.1 Description:**

Reusable Rack for Micro Centrifuge Tube, PC

#### **1.2 Design:**

Compatible Rack for 1.5 ml and 2.0 ml Micro Centrifuge Tube

#### **1.3 Performance Specifications:**

P20201 (1.5 & 2.0 ml), P20202 (1.5 & 2.0 ml), P20203 (1.5 & 2.0 ml), P20204 (1.5 & 2.0 ml)

#### **1.4 Material:**

1.4.1 Polycarbonate (PC), Confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests. Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE, 2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

**1.5 Color:** White & Yellow

#### **1.6 Graduations:**

Suitable for 1.5 & 2.0 ml Micro Centrifuge Tube

#### **1.7 Sterilization:**

Can be autoclaved at 15 Psi, 121 °C for 20 minutes



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

Registered office  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
Corporate Office  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email:labtech@abdosindia.com  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

## 1.8 Quality control:

Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

## 1.9 Other Information:

Reusable, can be autoclaved upto to 10 autoclave cycle.

## 2.0 Basic Features

### 2.1 Temperature range:

**2.2 Autoclavability:** Yes, Autoclavable at 15 psi and 121<sup>0</sup>C for 20 minutes

**2.3** Manufactured in Aseptic environment

**2.4 Chemical resistance:** See website see website ([www.abdoslifesciences.com](http://www.abdoslifesciences.com))

### Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121<sup>0</sup>C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50<sup>0</sup>C before removing product.

### Instruction for Use

Product can be used for food contact also.

Storage temp. Max 30<sup>0</sup>C

Before using the product check crackness, pouch punctured etc.

Check all the information printed on pouches or boxes.

Always wear gloves while handling specimen.

### WARNING

**DO NOT** store strong oxidizing agents.

**DO NOT** place any plastic labware in a flame.

**DO NOT** mix any chemicals that may result in a thermic reaction, which can cause product failure

Consult local fire codes prior to storage of flammable liquids.



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakheswari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

Registered office  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903  
Corporate Office  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email: [labtech@abdosindia.com](mailto:labtech@abdosindia.com)  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7

# Technical Data Sheet

## UKRP

Obelis UK LTD  
Samford Gate,  
East Point Business Park,  
Oxford OX4 5LH  
United Kingdom  
Tel: +44 1491 878012  
info@obelis.co.uk



EC REP

Obelis s.a.  
Rd General Wobis 53  
1030 Brussels Belgium  
Tel: +32 2 732-55-54  
Fax: +32 2 732-60-03  
mail@obelis.net




## Abdos Labtech Pvt. Ltd.

Plot No: 1 Lakshwari Bhagwanpur Roorkee 247667 (UK) India  
Email : labtech@abdosindia.com, www.abdoslifesciences.com  
Ph: +91 11 41081897



  
Quality Assurance Incharge  
Abdos Labtech Pvt. Ltd.



Abdos Labtech Private Limited  
(A Division of Abdos Group)  
(An ISO 9001,13485,14001,18385,GMP & CE certified company)  
Shiv Ganga Industrial Estate  
Plot No-1, Lakshwari, Bhagwanpur,  
Roorkee-247667, (UK) INDIA

Registered office  
13A Government  
Place East, 2nd Floor Kolkata  
West Bengal – 700069  
CIN: U85100WB2008PTC123903

Corporate Office  
319, 3rd Floor, DLF Tower A, DDA Dist. Centre  
Jasola, New Delhi - 110025

Ph. No: +91- (0) 1332235277  
Email:labtech@abdosindia.com  
Web: [www.abdoslifesciences.com](http://www.abdoslifesciences.com)  
GST: 05AAGCA8605A1Z7